



Picture shows similar calibration kit

Contents

Device	Part number	Quantity	Calibration Option ^a
Open circuit plug	51S12L-000S3	1	FC
Open circuit jack	51K12L-000S3	1	FC
Short circuit plug	51S12S-000S3	1	FC
Short circuit jack	51K12S-000S3	1	FC
Calibration load plug	51S170-C10S3	1	FC
Calibration load jack	51K170-C10S3	1	FC
Calibration adaptor plug/plug	51S121-S20S3	1	FC
Calibration adaptor jack/jack	51K121-K20S3	1	FC

a. See "Declaration of calibration options" for explanation.

Documentation

This kit is delivered with

- **USB-Stick**
Standard Definitions as data files for Vector Network Analyzer Families PNA (Keysight/Agilent) and ZVA (Rohde&Schwarz). Calibration Certificate as PDF-file.
- **Standard Definitions Cards**
Printed Standard Definitions that can be used on nearly all Vector Network Analyzers.
- **Kit Info Card**
Handling precautions and information for installing Standard Definitions on a Vector Network Analyzer.
- **Calibration Certificate**
Details see "Declaration of calibration options"
- **Operating Manual**

Electrical specifications

This specification covers electrical key values for the main calibration standards of the calibration kit. Specific datasheets are available for each component among the part number.

Calibration standard	Frequency	Parameter	Specification
Opens^b (plug and jack)	DC to ≤ 2 GHz > 2 GHz to ≤ 4 GHz	Error from Nominal Phase	≤ 3.0° ≤ 5.0°
Shorts^b (plug and jack)	DC to ≤ 2 GHz > 2 GHz to ≤ 4 GHz	Error from Nominal Phase	≤ 2.5° ≤ 4.0°
Calibration loads (plug and jack)	DC to ≤ 2 GHz > 2 GHz to ≤ 4 GHz	Return Loss	≥ 34 dB ≥ 30 dB
Calibration adaptors (plug/plug and jack/jack)	DC to ≤ 2 GHz > 2 GHz to ≤ 4 GHz	Return Loss	≥ 34 dB ≥ 30 dB

b. The specifications for opens and shorts are given as allowed deviation from nominal model as defined in calibration certificate included with your kit.

BNC
50 Ω

Calibration Kit
Industrial Version

51CK010-150

Declaration of calibration options

Factory Calibration

Standard delivery for this kit includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, **traceable to Rosenberger standards**, national / international standards are not available. Model based standard definitions of the calibration standards are reported in Agilent/Keysight, Rohde & Schwarz and Anritsu compatible VNA format.

Accredited Calibration

Not available.

For further, more detailed information see application note AN001 on the Rosenberger homepage.

Calibration interval

Recommendation 12 months

Recommended accessories

- Rosenberger Calibration Adaptor RPC-N female to BNC female 05K151-K20S3
- Rosenberger Calibration Adaptor RPC-N male to BNC male 05S151-S20S3
- Rosenberger Test Port Adaptor
- Rosenberger VNA Test cable kit and Microwave Cable Assemblies

For further, more detailed information please visit our homepage www.rosenberger.com.

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Herbert Babinger	05.05.15	Martin Moder	19.05.15	a00	15-s204	Maik Knoll	19.05.15
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel. : +49 8684 18-0 Fax : +49 8684 18-499 Email : info@rosenberger.de		Page 3 / 3